



**OFFICE OF THE MISSION DIRECTOR
NATIONAL HEALTH MISSION, ASSAM**

SAIKIA COMMERCIAL COMPLEX, CHRISTAN BASTI, G.S ROAD, GUWAHATI-781005

PH.NO : 0361-2363062 ; TELE FAX : 0361-2363058

Website: www.nrhmassam.in

e-mail: misnrhm.assam@gmail.com

Tender No: NHM/PROC/RBSK/3324/2017-18/18868

Date:9/10/2017

CORRIGENDUM No.2

TENDER FOR ENTERING INTO RATE CONTRACT OF COCHLEAR IMPLANT DEVICE AND DIGITAL PROGRAMMABLE HEARING AID

This has reference to the e-tender No: NHM Tender No: NHM/PROC/RBSK/3324/2017-18/10978 Date: 10/08/2017 for entering into rate contract for "Supply of Cochlear Implant Device and Digital Programmable Hearing Aid". The following clarification may be taken note of prior to submission of Bids:

1. The Cochlear Implant Device should have valid CE / USFDA / Drug Controller General of India (DCGI) Certificate.
2. For the Digital Programmable Hearing Aid Device,
 - a) The product should be CE/USFDA/IS 16127:2013 certified.
 - b) Maximum OSPL 90 should be 105 dB to 125 dB.
 - c) HF Average OSPL 90 should be 125 dB.
 - d) HF Average Full on Gain at 50/60 dB input should remain same as min 60 dB.
 - e) Microphone should be omni-directional.
 - f) The number of HD processing Channels in the Hearing Aid device should be 16 Channel or above.

All other terms and conditions of the tender referred to above shall remain unchanged. The specification provided in the tender referred to above are minimum. However item with higher technical specification will also be accepted.

Sd/-

Mission Director, NHM, Assam