



**OFFICE OF THE MISSION DIRECTOR
NATIONAL HEALTH MISSION, ASSAM
SAIKIA COMMERCIAL COMPLEX, CHRISTAN BASTI, G.S ROAD, GUWAHATI-781005
PH. NO: 0361-2363062; TELE FAX : 0361-2363058**

Website: www.nrhmassam.in

e-mail: misnrhm.assam@gmail.com

No: NHM/PROC/Misc Equip-Inst/2832/ 2015-16/7646

Date: 29/07/2016

CORRIGENDUM 1

e-TENDER FOR RATE CONTRACT OF HOSPITAL EQUIPMENTS AND INSTRUMENTS

This has reference to the e-tender No: NHM/PROC/Misc Equip-Inst/2832/ 2015-16/7121 Date: 22/07/2016 for entering into rate contract for supply & installation of Hospital Equipments & Instruments. The following amendment should be taken note of prior to submission of bid.

1. Earnest Money Deposit will remain fixed as Rs.2 Lakhs (Rupees Two Lakhs) only.
2. **Annexure-XII** (Technical cum Compliance Specification Sheet) of the tender is amended to read as follows:

Annexure XII (Rev-1)

Technical cum Compliance Specification Sheet: (TO BE SUBMITTED BY THE MANUFACTURER)

Sl. No	Tender Specification	Technical Specification quoted by the bidder	Deviation (if any)
1.	<p><u>Neonatal Resuscitation Kit :-</u></p> <ul style="list-style-type: none">• Infant Resuscitator with 250 ml Silicone Bellow, pressure release valve.• Silicone Face Mask (Anatomical): Autoclavable& reusable in sizes 00, 01• Foot Suction: Compact, lightweight, easy to handle and operate,• Durable rubber bellow,• Long lasting stainless spring to provide friction free pumping• Completely Autoclavable polycarbonate vacuum jar with lid (500 ml capacity), Scratch Proof power coated frame• Laryngoscope: Stainless Steel Straight blades of Size 0& 1 No.• Silicone airways: Autoclavable and reusable in Size 000& 00 No. <p>Should have CE,ISO 9001:2008& 13485:2003 certificates</p> <p>Warranty for 1yrs</p>		
2.	<p><u>Needle Destroyer-cum-Hub Cutter :-</u></p> <ul style="list-style-type: none">• Compact Device that fits easily on the desktop.		

Sl. No	Tender Specification	Technical Specification quoted by the bidder	Deviation (if any)
	<ul style="list-style-type: none"> • Portable, safe and cost effective system which creates save environment for Doctors, Nurse, Pathologist, Healthcare professionals and patients. • Cutter: Very high grade imported steel which does not rust or becomes blunt even after long use. • Body should make up of ABS plastic which avoids rusting and electric shock. • Electrodes should be copper electrode Nickle plate. • Temperature: Burning deg. should be 160°C. • Steel Grade should be SS316L which can cut syringe hub and burn needle approx. 25000 pcs to 28000pcs and life is 2-years minimum. • Supplied with Removable discharge tray for easy disposal of needle ash and syringe hubs. • Power consumption of continuous rating in use should be 60 Wtt. • Shock proof, fuse protection having, OFF & ON switch. <p>Should have CE, ISO 9001:2008 and ISO 13485:2003 certificates.</p> <p>Warranty for 2yrs</p>		
3.	<p>Foot Suction :-</p> <ul style="list-style-type: none"> • Max negative pressure valve : > 0.08MPa • Negative Pressure Back-reduction while being closed <0.005MPa/10 min • Reservoir Capacity :- 1000 ml • Pedal Pressure : < 245 N <p>Warranty for 2yrs</p> <p>Should have CE, ISO 9001:2008 and ISO 13485:2003 certificates.</p>		
4.	<p><u>Instruments Pack for Normal and Epistomy Delivery (comprising of Vacuum Extractor, Instruments and Catheter)</u></p> <ul style="list-style-type: none"> • Vacuum extractor: <ul style="list-style-type: none"> Vacuum Extractor Set , Bird Type, manual operated. Complete with 3 stainless steel cups of sizes 40mm, 50mm, 60 mm. ➤ Air Flow rate should be 36-40Ltr./min ➤ Autoclavable Silicone pressure tube. 		

Sl. No	Tender Specification	Technical Specification quoted by the bidder	Deviation (if any)
	<ul style="list-style-type: none"> ➤ Traction handle. ➤ Suction pump with suction rate (minimum-625mmHg or better) ➤ Glass jar fitted between M.S. Basket with vacuum gauge and control valve. ➤ 1 complete set in a carrying case. • Forceps obstetric, Wrigley's, 280 mm, stainless steel: - 1 Pcs. • Forceps, obstetric, Barnes-Neville, with traction, 390mm: - 1 Pcs. • Forceps, sponge holding, straight 228 mm, stainless steel: - 4 Pcs. • Forceps, artery, Spencer-Wells, straight, 180mm-ss: - 2 Pcs. • Forceps, artery, Spencer-Wells, straight, 140mm-ss: - 2 Pcs. • Holder, needle straight, Mayo-Hegar, 175 mm-ss: - 1 Pcs. • Scissors, ligature, Spencer, 130 mm, stainless steel: - 1 Pcs. • Scissors, episiotomy, angular, Braun, 145 mm, stainless steel: - 1 Pcs. • Forceps, tissue, spring-type, 1 x 2 teeth, 160 mm-ss: - 1 Pcs. • Forceps, tissue, spring-type, serrated ups, 160mm-ss :- 1 Pcs. • Catheter, urethral, rubber, Foley's 14 ER: - 1 Pcs. • Catheter, urethral, Nelaton, set of five (Fr 12-20) rubber: - 1 Pcs. • Forceps, backhaus towel -130 mm-ss: - 4 Pcs. • Speculum, vaginal, Sim's, double-ended # 3-ss:- 1 Pcs. • Speculum, vaginal, Hamilton-Bailey: - 1 Pcs. <p>Warranty for 2 yrs</p> <p>Should be CE and ISO 9001:2008 /13485:2003 certified.</p>		
5.	<p>Nitrous Oxide Cylinder 5 Ltrs:-</p> <ul style="list-style-type: none"> • Brand New High Pressure Seamless Cylinders for Nitrous Oxide Gas, duly filled with gas, 'A' type. • Having water Capacity 5Litres, Cylinders should be "ISI" Marked, Conforming to IS: 7285 Specification, • Fitted with Pin Index flush type Valve, painted as specified under Gas Cylinders Rules 1981. <p>Should have CE, valid ISI Certificate and ISO</p>		

Sl. No	Tender Specification	Technical Specification quoted by the bidder	Deviation (if any)
	9001:2008/13485:2003 Certificate. Warranty for 2yrs		
6.	<u>Oxygen Cylinder A Type:-</u> <ul style="list-style-type: none"> • Brand New High Pressure Seamless Cylinders for Oxygen Gas, duly filled with gas 'A' type. • Having water Capacity 5Litres, Cylinders should be "ISI" Marked, Conforming to IS: 7285 Specification, • Fitted with Pin Index flush type Valve, painted as specified under gas cylinders rules 1981. Should have CE, valid ISI Certificate and ISO 9001:2008/13485:2003 Certified Warranty for 2yrs		
7.	<u>Instruments pack for C-Section and other operations</u> (Comprising of (a). Delivery Kit, (b). Epistomy kit and c. Forceps Delivery Kit) <ul style="list-style-type: none"> ○ <u>Delivery Kit :-</u> <ul style="list-style-type: none"> • Trolley, dressing carriage <ul style="list-style-type: none"> ➤ Size 76C, long x 46 cm wide and 84 cm high. Ref. IS 4769/1968 :- 1Pc. ➤ Towel Trolley 84 cm x 54 cm :- 2Pc. • Tray instrument with cover, Size: 450 mm (L) x 300 mm (W) x 80 mm (H) :- 1 Pc • Drum, sterilizing cylindrical, Size – 275 mm Dia x 132 mm, ss as per IS: 3831/1979 :- 02 Pc • Table instrument adjustable type with tray ss :- 1Pc. ○ <u>Episiotomy kit :-</u> Should consist of the following instruments :- <ul style="list-style-type: none"> • Vacuum extractor: <ul style="list-style-type: none"> Vacuum Extractor Set, Bird Type, manual operated. Complete with 3 stainless steel cups of sizes 40mm, 50mm, 60 mm. ➤ Air Flow rate should be 36-40Ltr./min ➤ Autoclavable Silicone pressure tube. ➤ Traction handle. ➤ Suction pump with suction rate (minimum-625mmHg or better) 		

Sl. No	Tender Specification	Technical Specification quoted by the bidder	Deviation (if any)
	<ul style="list-style-type: none"> ➤ Glass jar fitted between M.S. Basket with vacuum gauge and control valve. ➤ 1 complete set in a carrying case. • Forceps obstetric, Wrigley's, 280 mm, stainless steel :- 1 Pc • Forceps, obstetric, Barnes-Neville, with traction, 390mm :- 1 Pcs • Forceps, sponge holding, straight 228 mm, stainless steel :- 4 Pcs • Forceps, artery, Spencer-Wells, straight, 180mm-ss :- 2 Pcs • Forceps, artery, Spencer-Wells, straight, 140mm-ss :- 2 Pcs • Holder, needle straight, Mayo-Hegar, 175 mm-ss :- 1 Pc • Scissors, ligature, Spencer, 130 mm, stainless steel :- 1 Pc • Scissors, episiotomy, angular, Braun, 145 mm, stainless steel :- 1 Pc • Forceps, tissue, spring-type, 1 x 2 teeth, 160 mm-ss :- 1 Pc • Forceps, tissue, spring-type, serrated ups, 160mm-ss :- 1 Pc • Catheter, urethral, rubber, Foley's 14 ER :- 1 Pc • Catheter, urethral, Nelaton, set of five (Fr 12-20) rubber :- 1 Pc • Forceps, backhaus towel -130 mm-ss :- 4 Pcs • Speculum, vaginal, Sim's, double-ended # 3-ss :- 1 Pc • Speculum, vaginal, Hamilton-Bailey :- 1 Pc <p>A. <u>Forceps Delivery Kit</u> :-</p> <p>Should consist of the following instruments :-</p> <ul style="list-style-type: none"> • Forceps, obstetric, Neville-Barnes, W/traction 390 mm :- 1 Pc • Hook, decapitation, Braun, 300 mm, stainless steel :- 1 Pc • Hook, crochet, obstetric 300 mm, Smellie, stainless steel :- 1 Pc • Bone, forceps, Mesnard 280 mm, stainless steel :- 4 Pcs • Perforator, Smellie, 250 mm, stainless steel :- 1 Pc • Forceps, cranial, Gouss, straight, 295 mm-ss :- 1 Pc • Cranioclast, Braun, stainless steel, 365 mm long :- 1 		

Sl. No	Tender Specification	Technical Specification quoted by the bidder	Deviation (if any)
	<p>Pc</p> <ul style="list-style-type: none"> • Scissors ligature Spencer 130 mm, stainless steel :- 1 Pc • Forceps sponge holding, 22.5 cm straight – ss :- 1 Pc • Forceps, tissue, spring-type, 1 x 2 teeth, 160 mm, stainless steel :- 1 Pc • Forceps, tissue, spring-type, serrated tips, 160 mm-ss :- 1 Pc • Forceps, artery, Spencer–Wells, straight, 180 mm-ss :- 2 Pcs • Forceps, artery, Spencer-Wells, straight, 140 mm-ss :- 2 Pcs • Forceps, scalp flap, Willet’s 190 mm –ss :- 4 Pcs • Forceps, Vulsellum, duplay double curved, 280 mm-ss :- 4 Pcs • Forceps, Vulsellum, duplay double curved, 240 mm-ss :- 01 Nos. • Catheter, urethral, 14 Fr. Solid tip, one eye,-: 3 Pcs. • Holder, needle, Mayo-Hegar, narrow jaw, straight, 175 mm-ss: - 01 Nos. • Speculum vaginal bi-valve, Cusco-medium, stainless steel: - 01 Nos. • Speculum, vaginal sim’s double-ended, size # 3-SS: - 01 Nos. • Forceps, Backhaustowel, 130 mm, stainless steel: - 01 Nos. <ul style="list-style-type: none"> • All instruments of this kit should be of medical grade stainless steel. • Should be CE and have valid ISI Certificate, ISO 9001:2008/13485:2003 • Warranty for 2yrs 		
8.	<p><u>Kidney Tray SS.:-</u></p> <ul style="list-style-type: none"> • Kidney Tray SS.As Per IS: 3992-1982 (Heavy Quality) • Size: 250 x 100 • Should be ISO 9001:2008 Certified and have valid ISI Certificate. • Warranty for 2yrs 		
9.	<p><u>Instrument Tray with cover SS.:- (304 Gauge 18)</u></p> <ul style="list-style-type: none"> • Instrument Tray with cover SS Seamless IS 3993-192.The finished thickness should not be less than 1.2 mm. 		

Sl. No	Tender Specification	Technical Specification quoted by the bidder	Deviation (if any)
	<ul style="list-style-type: none"> • Should be ISO 9001:2008 Certified and have valid ISI Certificate. • Size: 350 x 250 mm • Warranty for 2yrs 		
10.	<p><u>Dressing Tray without cover SS:</u></p> <ul style="list-style-type: none"> • Dressing Tray SS (304 Gauge 18) • Size: 300x250 in mm (12"x10") • Should be ISO 9001:2008 Certified • Warranty for 2yrs 		
11.	<p><u>X-Ray View Box :-</u></p> <ul style="list-style-type: none"> • X-Ray View Box Double Film with two tubes and epoxy powder coated deluxe. • Should be CE, ISO 9001:2008& ISO 13485:2003 Certified • Warranty for 2 yrs 		
12.	<p><u>Vertical Autoclave, 30 ltrs :-</u></p> <ul style="list-style-type: none"> • Electrical Portable vertical sterilizer. • Autoclave seamless is made out of thick high quality stainless steel sheet, deep drawn to cylindrical shape, having no joints and welding, seamless construction allows no bacteria residue and free from dirt accumulation with tripod stand made from MS and Epoxy Powder Coated. • Size: 350 mm diameter x 300-325 mm height. • Should be ISI Marked. • Should be CE, valid ISI and ISO 9001:2008/13485:2003 Certified <p>Spare Parts:</p> <ol style="list-style-type: none"> 1. Heating Element 2 KW: - 01 Pcs. 2. Wire with Top & Connector: - 01 Pcs. 3. Rubber Gasket: - 02 Pcs. 4. Dressing Drums Seamless ISI Marked, Size: 225x225 mm : - 01 Pcs. 5. Dressing Drums Seamless ISI Marked, Size: 275x240 mm : - 01 Pcs. <p>Warranty:-02 years</p>		

Sl. No	Tender Specification	Technical Specification quoted by the bidder	Deviation (if any)
13.	<p><u>Stethoscope: Adult:</u></p> <ul style="list-style-type: none"> • Adult size chest piece Precisely Machined to extracting tolerance from surgical stainless steel for outstanding performance and rugged durability. • Extra Large bell for unsurpassed low frequency response. • Ultra –Sensitive Diaphragm for greater amplification • Colour coordinated Non-Chill bell and Snap on Ring to retain diaphragm for patient comfort. • Deluxe PVC ear tips for the ultimate in wearing comfort and acoustic seal • Spare Diaphragm for patient comfort • Including ID Tag • In foam protective packaging to provide safety to the stethoscope • Minimum length should be 60 cms. • Should be ISO 9001:2008 certified. • Warranty: - 01 year. 		
14	<p><u>Stethoscope:Neonatal Stethoscope:</u></p> <ul style="list-style-type: none"> • Neonatal size chest piece Precisely Machined to extracting tolerance from surgical stainless steel for outstanding performance and rugged durability. • Extra Large bell for unsurpassed low frequency response. • Ultra –Sensitive Diaphragm for greater amplification • Colour coordinated Non-Chill bell and Snap on Ring to retain diaphragm for patient comfort. • Deluxe PVC ear tips for the ultimate in wearing comfort and acoustic seal • Spare Diaphragm for patient comfort • Including ID Tag • In foam protective packaging to provide safety to the stethoscope • Minimum length should be 60 cms. • Warranty:-01 year. • Should be ISO 9001:2008Certified. 		
15.	<p>Urinal (Male)</p> <ul style="list-style-type: none"> • Should be made from 304 grade Stainless Steel, capacity 700 ml minimum. 		

Sl. No	Tender Specification	Technical Specification quoted by the bidder	Deviation (if any)
	Should be ISO 9001:2008 certified. Warranty: - 01 year.		
16.	Urinal (Female) <ul style="list-style-type: none"> • Should be made from 304 grade Stainless Steel, capacity 700 ml minimum. • Should be ISO 9001:2008 certified. • Warranty: - 01 year. 		
17	Bed Pan(Female) <ul style="list-style-type: none"> • Leak Proof, Stainless Steel (304 Gauge 18) with Argon welding. • Capacity:1 L Warranty:1 year Should be ISO 9001:2008 certified.		
18	Bed Pan(Male) <ul style="list-style-type: none"> • Leak Proof, Stainless Steel (304 Gauge 18) with Argon welding. • Capacity:1 L Warranty:1 year Should be ISO 9001:2008 certified.		

All other terms and conditions of the tender referred to above shall remain unchanged.

Sd/-
Mission Director, NHM, Assam