



OFFICE OF THE MISSION DIRECTOR
NATIONAL HEALTH MISSION, ASSAM
SAIKIA COMMERCIAL COMPLEX, CHRISTAN BASTI, G.S ROAD, GUWAHATI-781005
PH.NO: 0361-2363062; TELE FAX: 0361-2363058

Website: www.nrhmassam.in

e-mail: misnrhm.assam@gmail.com

No: NHM/PROC/CH/Training Mannequin/3067 /2016-17/18154

Date: 09/01/2017

CORRIGENDUM 1

e- TENDER FOR SUPPLY OF TRAINING MANNEQUINS

This has reference to the e-tender No: Tender No.: NHM/PROC/CH/Training Mannequin/3067 /2016-17/15971 Date: 19/12/2016 for "SUPPLY OF TRAINING MANNEQUINS" to different Health Institutions in the state. The following amendment should be taken note of prior to submission of bid.

1. Clause 5 (Point No: J) Annual Turnover of the Bidder is amended to read as follows:

Annual turnover statement from sales of Manikins, Simulators & Model to Medical instruments for last three financial years i.e. 2013-14, 2014-15 & 2015-16 in the format given in **Annexure-III A & B** certified by the Auditor/Chartered Accountant. It is to be noted that such statement of both the bidder and the manufacturer (in case the bidder is not a manufacturer) must be submitted.

Certificate of Auditor/Chartered Accountant must be accompanied by attested/notarized copies of audited "Profit & Loss Statement" and "Balance Sheet" of last three financial years (2013-14, 2014-15 & 2015-16). These statements including "Annual Turnover Statement", "Profit & Loss Statement" and "Balance Sheet" will not be required for overseas manufacturers. However such statements of bidders need to be submitted.

2. Annexure XII "Technical Specification cum Compliance Statement" is amended to read as follows:

Annexure XI (Revised-1)

**Technical Specification cum Compliance Statement:
(TO BE SUBMITTED BY THE MANUFACTURER)**

Sl. No	Name of items	Technical Specifications	Quantity required	Technical Specs quoted by bidder	Deviation if any
1.	Neonatal Mannequin for demonstrating Essential Newborn Care and Resuscitation	<ul style="list-style-type: none"> • Should be Air/Water Filled. • Presence of anatomical landmarks. • Manual simulation for breathing with adequate chest rise and depression. • Manual simulation for suctioning techniques. • Positive pressure ventilation simulation with AMBU Bag. • Able to perform Oxygen airway procedures. • Provision to demonstrate Chest Compressions. • Umbilical cord with provision of manual simulation of pulse. • Simulated crying sounds. • Mannequin must be supplied along with the Bag and Mask set having 0 & 1 size mask. • Size & Weight to exactly match as per the dimensions of a full term newborn delivered between 39 wks 0 days & 40 wks 6 days <p><u>Warranty: 3 years</u> Quarterly visit to site, within warranty period including training of user on maintenance.</p> <p><u>Certifications</u> BS EN ISO 13485:2003/IEC 17050-1:2010</p> <p><u>Documentation</u></p> <ul style="list-style-type: none"> • Operating manuals, service manuals, other manuals • Advanced maintenance tasks required shall be documented • User manuals to be supplied in English/Hindi language along with machine diagrams. 	200 nos.		

2.	Advanced Mannequin for demonstration of Neonatal Resuscitation	<ul style="list-style-type: none"> • Colour change with conditions / interventions (Automatic simulation to depict Cyanosis & its interventions) • Airway to simulate intubation. • Automatic simulation of lung sounds. • Simulator for respiration, rate, depth & chest rise. • Bilateral IV arms with fill/drain sites. • Umbilical cord with provision for catheterisation & pulse rate simulation. • Programmed simulation to depict different conditions for neonatal resuscitation related procedures. • Size & Weight to exactly match as per the dimensions of a full term newborn delivered between 39 wks 0 days & 40 wks 6 days • Skin tone matching with local population. <p><u>Warranty: 3 years</u> Quarterly visit to site, within warranty period including training of user on maintenance.</p> <p><u>Certifications</u></p> <p>BS EN ISO 13485:2003/IEC 17050-1:2010</p> <p><u>Documentation</u></p> <ul style="list-style-type: none"> • Operating manuals, service manuals, other manuals • Advanced maintenance tasks required shall be documented • User manuals to be supplied in English/Hindi language along with machine diagrams. 	8 nos.		
----	---	--	---------------	--	--

The bid submission start date and end date has been extended to 25/01/2017 & 03/02/2017 respectively.

All other terms and conditions of the tender referred to above shall remain unchanged.

Sd/-
Mission Director, NHM, Assam